

INTENDED USE

The Detectabuse controls are designed to monitor and validate the performance of drugs of abuse detection methods at levels established by SAMHSA, CAP/AACC and many state programs. The Detectabuse control urines are compatible with all quantitative and qualitative drug detection procedures which are sufficiently sensitive to detect the control constituents. They should be treated as any "unknown" specimen while following the specific protocol of the assay being used.

This product is intended to be used by health care professionals as an integral part of good laboratory practices.

SUMMARY AND EXPLANATION

The DEA exempt Detectabuse product line of controls is manufactured using a human based matrix that has been stabilized to insure that the product will be viable until the date of expiration. Positive controls are spiked with reference drug standards and/or appropriate metabolites that have been obtained from ISO certified manufacturers. Standards are certified by the manufacturer to be at least 98% minimum purity. Specific gravity, pH, and creatinine fall within the limits of normal human urine.

DESCRIPTION

Each bottle contains stabilized human based urine. Positive control urines have been spiked with authentic reference drug standards and/or appropriate metabolites. Negative control urines are certified negative by a combination of immunoassay, GC/MS and/or LC/MS for the constituents listed on our target sheets.

CATALOG #	DESCRIPTION	
1933200-1	GC/MS Cutoff -25% (Low Opiate)	50 mL
1933000-0	GC/MS Cutoff (Low Opiate)	50 mL
1933250-1	GC/MS Cutoff +25% (Low Opiate)	50 mL
1957225-1	GC/MS-H Cutoff -25% (High Opiate)	50 mL
1957250-2	GC/MS-H Cutoff (High Opiate)	50 mL
1957275-3	GC/MS-H Cutoff+25% (High Opiate)	50 mL
1958200-1	Confirm 2, Cutoff -60% (High Opiate)	50 mL
1958222-5	Confirm 2, Cutoff (High Opiate)	50 mL
1958250-1	Confirm 2, Cutoff +25% (High Opiate)	50 mL
1922700-0	Detectabuse Liquid Control Urine, Negative	50 mL

**Call for Confirm2, Cutoff -25% (High Opiate) information

For additional information on our other Detectabuse products please contact us or refer to our website: www.biochemicaldiagnostics.com

PRECAUTIONS

For in vitro diagnostic use only

Please read the entire package insert before using the Detectabuse control urines. Please use the same safety precautions you would use for processing any "unknown" urine sample containing potentially infectious biological material. Protect product from exposure to direct sunlight.

Contains sodium azide: To prevent formation of explosive metal azides dispose of waste by flushing with copious amounts of water or according to local governing regulations.

Do not use beyond the expiration date.

STORAGE & STABILITY - Please refer to Technical Note for detailed instructions.

Unopened:

- The controls are stable until the expiration date when stored at -10 to -20°C and protected from light.
- The controls are stable until the expiration date when stored at 2-8°C, however, no stability claims can be made for Oxazepam and/or 6-MAM as they may deteriorate over time when stored refrigerated.

After Opening:

- The controls are stable for six months or until the expiration date, whichever comes first, when stored at -10 to -20°C. (Controls can be aliquoted and frozen)
- The controls are stable for 31 days or until the expiration date, whichever comes first, when stored tightly capped at 2-8°C.
- Thaw controls as needed; allow to come to room temperature followed by gentle swirling before use.

PROCEDURE

- Allow controls to come to room temperature followed by gentle swirling or inversion before use. **DO NOT SHAKE.**
- Pipette an appropriate aliquot of Detectabuse control urine as required by the confirmation method.

Detectabuse® GC/MS Liquid Control Urine

Target Values (ng/mL)

SAMHSA MANDATED	Cutoff -25%	Cutoff	Cutoff +25%
Delta-9-THC-COOH	11	15	19
Benzoylcegonine	113	150	188
Phencyclidine (PCP)	19	25	31
Codeine (Low Opiate)	225	300	375
Codeine (High Opiate)	1500	2000	2500
*Total Morphine (Low Opiate)	225	300	375
* Total Morphine (High Opiate)	1500	2000	2500
**6-Monoacetylmorphine	7.5	10	12.5
d-Amphetamine	375	500	625
d-Methamphetamine	375	500	625
NON-MANDATED			
Secobarbital	225	300	375
Phenobarbital	225	300	375
Oxazepam	225	300	375
Methadone	225	300	375
Methaqualone	225	300	375
Propoxyphene	225	300	375
Butalbital	225	300	375

Detectabuse® CONFIRM 2 Liquid Control Urine

Target Values (ng/mL)

SAMHSA MANDATED	Cutoff -60%	Cutoff	Cutoff +25%
Delta-9-THC-COOH	6	15	18.75
Benzoylcegonine	40	100	125
Codeine	800	2000	2500
*Total Morphine	800	2000	2500
6-Monoacetylmorphine	4	10	12.5
Phencyclidine (PCP)	10	25	31.25
d-Methamphetamine	100	250	312.5
d-Amphetamine	100	250	312.5
MDMA	100	250	312.5
MDA	100	250	312.5
MDEA	100	250	312.5

Negative Control – Negative for the constituents listed on our target sheets.

* Contains Morphine-3-Glucuronide and Free Morphine (2:1)

** 6-MAM Only in High Opiate (-H) Control

TECHNICAL NOTE:

DETECTABUSE CONTROLS, THC STABILITY

Detectabuse controls are stable for the length of time under the storage conditions stated in the package insert. In spite of this fact, under certain conditions, there may be observed a gradual decline in THC levels, over time, from continuous use of a single bottle of control material. This drop in THC values may occur from any THC sample (i.e. calibrators, controls, and samples).

The apparent loss of THC most often occurs from handling and not from product instability. It is well known that THC-COOH binds to surfaces, especially certain plastics. In order to minimize this adsorption loss we recommend the following when handling any sample (including Detectabuse controls) which may contain THC:

1. Preferably, use glass pipettes or pour controls into sample cups.
As an alternate, pipettors with disposable plastic tips may be used.
Soft plastic transfer pipettes should be avoided.
2. Do not rinse the pipette back and forth into the sample.
3. Sample volume to surface area ratio should be as high as possible (i.e. when transferring, sample containers should be filled as much as possible with sample). Avoid rough surface plastic containers.
4. When pipetting, immerse the pipette tip as little as possible into the sample solution.
5. Do not return any unused material back into the original sample.

These same guidelines should also be followed when aliquoting a control (or sample) for future use.

References: 1. Blanc JA, Manneh VA, et al. Adsorption losses from urine-based cannabinoid calibrators during routine use. Clin Chem 1993; 39:1705-1712

2. Roth KDW, Siegel NA, et al. Investigation of the effects of solution composition and container material type on the loss of 11-nor-delta 9-THC-9-carboxylic acid. J Anal Tox 1996; 20:291-300

DETECTABUSE CONTROLS, OXAZEPAM STABILITY

Oxazepam has known stability problems in urine stored refrigerated, and to a lesser degree, frozen. Our experience indicates that Oxazepam will not deteriorate more than 10% of target level for at least one year when stored frozen at -20°C. Further deteriorations may occur beyond this period although Oxazepam ordinarily tests positive throughout the control's shelf life.

LIMITATIONS OF PROCEDURE

This control is meant to be used to validate the performance of GC/MS confirmation methods. Consult test manufacturers instructions when using this product; changes in reagents, sample requirement, or methodology may effect test results.

Although target values are provided with the Detectabuse liquid controls, each laboratory should run these controls as unknowns in order to establish "in-house" assay values for them.

This product is not meant to be used as a standard or calibrator.

Expected Results

The positive Detectabuse control must test positive. The negative control must test negative. BiochemicalDiagnostics will (upon request), supply assay values derived from our contract assay laboratories and customer base on a particular lot of control material.



Biochemical
Diagnostics, Inc.

DETECTABUSE®

LIQUID

CONTROL

URINE

GC/MS (Low Opiate)

GC/MS-H (High Opiate)

Target levels available: Negative, Cutoff, Cutoff -25%, Cutoff +25%

CONFIRM 2 (New SAMHSA guidelines)

Target levels available: Cutoff -60%, Cutoff, Cutoff +25%,

Controls prepared from human based urine available as a negative and at various constituent target levels to monitor the performance of qualitative or quantitative procedures for the detection of drugs in urine.

Please read the entire package insert before using the Detectabuse control urines.



Temperature Limit



Consult Instructions for Use



For In Vitro Diagnostic Use

180 Heartland Blvd, Edgewood, NY 11717 • Phone (800) 223-4835
Fax (631) 595-9204 • www.biochemicaldiagnostics.com